



CORRIGENDUM-I

**Ref :TENDER SPECIFICATION NO. SR.G.M.CPC-TENDER-TESTING EQUIPMENTS-29/2010-11,
:TENDER NOTICE NO. 23/2010-11.**

Package-V: 3Ph,4W Reference Standard Meter(HT+0.05 Acc Class):

The following Technical points for Portable Three Phase Reference Standard Meter of Accuracy class 0.05 Capable of testing of Energy Meters:-

- 1) The Reference standard Meter shall be of without C.T and the specification related to testing of LT C.T should be deleted.**
- 2) The Ratio error and phase error of C.T indicated in any where in the specification shall stand deleted.**
- 3) TS Clause 2.0 for"Application" shall be read as: The function of the portable Electronic Reference standard meter shall be suitable to measure the system parameters and verify the accuracy of three phase and single phase energy meters in the laboratory and at site.**
- 4) GUARANTEED TECHNICAL PARTICULARS OF UNIVERSAL, PORTABLE ELECTRONIC REFERENCE SUB STANDARD (ERSS) METER OF ACCURAY 0.05 IN DIRECT MODE FOR TESTING OF METERS.**

**All other Terms and condition shall be as per the Tender Specification No. Sr. G.M -CPC-TENDER-
TESTING EQUIPMENT – 29/2010-11.**